

NATIONAL COIR RESEARCH AND MANAGEMENT INSTITUTE (NCRMI)
KUDAPPANAKUNNU.P.O, THIRUVANANTHAPURAM

NCRMI/R&D/2002/16

05/07/2018

QUOTATION NOTICE

Competent quotations are invited for the supply and installation of Hot air oven in Geotechnical lab at NCRMI with following specification

Specification

Size : 605 x 605 x 605mm

Inside chamber : Aluminium / Stainless steel

Temperature range : +5⁰Cup to 250⁰C Thermostatic control with a sensitivity of $\pm 1^0C$

Certification : CE and ISO certified

Heating element placed in the ribs at the bottom and sides

Double walled –outside of mild steel powder coated with 4nos. of perforated adjustable shelves, thermostat, two indicators on/off switch work on 220/230V AC

The instrument should come with one circulating fan and electronic digital microprocessor based temperature controller cum indicator.

The interested dealers and parties may quote their prices along with the taxes payable. The quotations shall reach the Director, NCRMI on or before 19/07/2018 at 3.00PM. The quotations will be opened at 3.30PM on the same day in the presence of quotationers present. The Director, NCRMI reserves the right to accept or reject any or all of the quotations without assigning any reason whatsoever.

DIRECTOR
NCRMI

To

NCRMI Website & Notice board

