



NATIONAL COIR RESEARCH & MANAGEMENT INSTITUTE
(NCRMI)
(A Govt. of Kerala Undertaking)



NCRMI/R&D/3016/22

15.10.2022

QUOTATION NOTICE

Competitive and sealed quotations are invited from the suppliers for the supply of Lab scale Analytical Balance of following specifications:

Sl No	Description	Specifications	Qty
1	Analytical Balance	<ul style="list-style-type: none">• Capacity : 200g• Readability : 0.1mg• Repeatability : 0.1mg• Linearity(+/-) : 0.2mg• Response time : 2-3sec• Display : Alpha numeric backlite LCD display• Calibration : Motorised internal calibration• Units of measure : g, mg etc• Tare range : Full• Operating temperature: 15^o C to 35^o C• Sensitivity driff : ±2 ppm• Power supply : A/C adaptor 230V• Housing dimension : 212 × 341× 345• Warranty : 2 years	1

The quotations shall reach The Director, NCRMI on or before 25.10.2022 at 11.00 AM. The quotations will be opened at NCRMI in the presence of quotationers, if any present at 11.30 AM on the same day. The Director reserves the right to accept or reject all or any of the quotations without assigning any reasons whatsoever.

DIRECTOR

NCRMI